

ClarifyCRM - ClearSupport - [[READ] Case 0024180]

File Edit View Desktop New Select Actions Apps Window Help

[[READ] Case 0024180

Salutation: Title:

Contact Name (F/L): Monica Revello Main Phone: 5065805

Site ID/Name: 0051061352 DIR. NAC. SANIDAD POLICIAL

ZIP/City/Country: 12000 MONTEVIDEO Uruguay

SN/PN/Desc.: 92707 11568248001 Elecsys 2010 disk *IM

Family/Line: Immunology ELECSYS 2010

SA-Group/Code: Repair Visit ServVisit Assist

Case Type / ID: Service Report 0024180

Case Spec.Consid. Prev.Cases/Int. More Info Site/Contact Attributes Reag/Acc Instrument Description Scheduling Contract Billing T&E/Parts Log/Price Parent Case Child Cases

Billing Info

Expense Account:

FSE: Venturino, Horacio, +598-2-613 7888 Clear FSE

FSE SiteID: INT1 Log FSE activity Case FSE activity

Work Performed: Please Specify

Cust. Order No:

Cust. Order Date: 11/11/1

Effective Visit Date: 17/02/2014 23:00:00

First Contact Date: ?/?/?

Requested Date: ?/?/?

Change Bill-to Site... Bill-to Service

0051061352 DIR. NAC. SANIDAD POLICIAL

BATLLE Y ORDOÑEZ 3574

12000 MONTEVIDEO

Pay Terms: Please Specify

Ship Via: Please Specify

Invoice Notes (Problem Description):

Se dañó disquete y no inicia. Se coloca disquete de datos de 2012 de cliente y en la instalación del SW 6.06/4.05 se le presenta ese disquete para que tome el Adjust.dat. Con eso se logra que realice HV y BCinit. Colaboré con MH